



Material suplementario

Trombosis tardía asociada a dispositivo de cierre percutáneo de orejuela. ¿Es poco frecuente o está infradiagnosticada?

Tabla del material suplementario

Revisión bibliográfica de casos publicados de trombosis asociada a dispositivos de cierre percutáneo de orejuela

Artículo	Año publicación	Tipo de publicación	Dispositivo	Casos de trombosis	Diagnóstico de trombosis ^a				Resolución
					≤ 6 meses	> 6—≤ 12 meses	> 12—≤ 24 meses	> 24 meses	
Sick et al. ¹	2007	Registro multicéntrico (n = 66)	Watchman	4	4	—	—	—	Anticoagulación
Cardona et al. ²	2011	Comunicación de caso	ACP	1	1	—	—	—	Anticoagulación
Cruz-González et al. ³	2011	Comunicación de caso	ACP	1	1	—	—	—	Anticoagulación
Bai et al. ⁴	2012	Registro unicéntrico (n = 58)	Watchman	1	—	—	1 ^b	—	Anticoagulación
Chowdhury et al. ⁵	2012	Comunicación de caso	Watchman	1	1	—	—	—	Anticoagulación
Gasparini et al. ⁶	2012	Comunicación de caso	Watchman	1	1	—	—	—	Anticoagulación
Lammers et al. ⁷	2013	Comunicación de caso	ACP	1	1	—	—	—	Cirugía
Plicht et al. ⁸	2013	Registro unicéntrico (n = 34)	ACP	6	6	—	—	—	Anticoagulación
Elabbassi et al. ⁹	2013	Comunicación de caso	Watchman	1	1	—	—	—	Anticoagulación
Fernández-Rodríguez et al. ¹⁰	2013	Comunicación de caso	ACP	1	1	—	—	—	Anticoagulación
López-Mínguez et al. ¹¹	2013	Registro unicéntrico (n = 35)	ACP	5	5	—	—	—	Anticoagulación
Meincke et al. ¹²	2013	Registro unicéntrico (n = 59)	Watchman	3	3	—	—	—	Anticoagulación
Chun et al. ¹³	2013	Registro unicéntrico	Watchman (38) + ACP	4	4	—	—	—	Antiagregación o anticoagulación

		(n = 78)	(40)						
Faustino et al. ¹⁴	2013	Registro unicéntrico (n = 22)	Watchman (2) + ACP (20)	1	1	—	—	—	Anticoagulación
Dilling-Boer et al. ¹⁵	2014	Comunicación de caso	ACP	1	—	—	1	—	Cirugía
Schiettekatte et al. ¹⁶	2014	Comunicación de caso	ACP	1	—	—	1	—	Cirugía
Ternacle et al. ¹⁷	2014	Comunicación de caso	ACP	1	1	—	—	—	Anticoagulación
Wiebe et al. ¹⁸	2014	Registro unicéntrico (n = 57)	ACP	2	2	—	—	—	Anticoagulación
Freixa et al. ¹⁹	2014	Registro unicéntrico (n = 24)	Amulet	1	1	—	—	—	Anticoagulación
Gafoor et al. ²⁰	2014	Registro unicéntrico (n = 49)	Watchman (26) + ACP (27)	1	1	—	—	—	Anticoagulación
López-Mínguez et al. ²¹	2015	Registro multicéntrico (n = 158)	ACP	13	—	13 ^c	—	—	Anticoagulación
Ketterer et al. ²²	2015	Registro unicéntrico (n = 46)	Watchman	4	4	—	—	—	Anticoagulación
Kebernik et al. ²³	2015	Registro unicéntrico (n = 96)	ACP	2	2	—	—	—	Anticoagulación
Freixa et al. ²⁴	2015	Comunicación de caso	Amulet	1	—	1	—	—	Anticoagulación
Świątkiewicz et al. ²⁵	2016	Comunicación de caso	Watchman	1	1	—	—	—	Cirugía
Qazi et al. ²⁶	2016	Comunicación de caso	Watchman	1	—	—	1 ^d	—	Anticoagulación

Ledwoch et al. ²⁷	2016	Registro unicéntrico	Watchman	1	1	—	—	—	Anticoagulación
Saw et al. ²⁸	2017	Registro multicéntrico (n = 339)	ACP	11	—	—	11 ^c	—	Antiagregación o anticoagulación

ACP: Amplatzer Cardiac Plug; ETE: ecocardiografía transesofágica.

^aMomento en el que se detectó trombosis asociada a dispositivo (≤ 6 meses; > 6 y ≤ 12 meses; > 12 y ≤ 24 meses o > 24 meses tras implante del dispositivo).

^bRegistro unicéntrico de seguimiento de casos de oclusión incompleta de orejuela izquierda.

^cNo se detalla en la publicación el momento exacto del diagnóstico de trombosis.

^dRecurrencia de trombosis a los 24 meses diagnosticada en la ETE a los 12 meses.

BIBLIOGRAFÍA

1. Sick PB, Schuler G, Hauptmann KE, et al. Initial worldwide experience with the WATCHMAN left atrial appendage system for stroke prevention in atrial fibrillation. *J Am Coll Cardiol*. 2007;49:1490-1495.
2. Cardona L, Ana G, Luísa B, et al. Thrombus formation on a left atrial appendage closure device. *Circulation*. 2011;124:1595-1596.
3. Cruz-Gonzalez I, Martín Moreiras J, García E. Thrombus formation after left atrial appendage exclusion using an Amplatzer cardiac plug device. *Catheter Cardiovasc Interv*. 2011;78:970-973.
4. Bai R, Horton RP, Di Biase L, et al. Intraprocedural and longterm incomplete occlusion of the left atrial appendage following placement of the WATCHMAN device: A single center experience. *J Cardiovasc Electrophysiol*. 2012;23:455-461.
5. Chowdhury MA, Elabbassi W, Al Nooryani A. Left atrial appendage occlusion device-related thrombus: a Comunicación de caso and review of the literature. *J Echocardiogr*. 2012;10:65-66.
6. Gasparini M, Ceriotti C, Bragato R. Huge left atrial thrombus after left atrial appendage occlusion with a Watchman device. *Eur Heart J*. 2012;33:1998.
7. Lammers J, Elenbaas T, Meijer A. Thrombus formation on an Amplatzer closure device after left atrial appendage closure. *Eur Heart J*. 2013;34:741.
8. Plicht B, Konorza TF, Kahlert P, et al. Risk factors for thrombus formation on the Amplatzer Cardiac Plug after left atrial appendage occlusion. *JACC Cardiovasc Interv*. 2013;6:606-613.
9. Elabbassi W, Osman AN, Chowdhury MA, et al. Device-related thrombus overlying a left-atrial occlusion device implanted for persistent spontaneous echo contrast despite prolonged oral anticoagulation therapy. *Cardiology*. 2013;125:78-81.
10. Fernández-Rodríguez D, Vannini L, Martín-Yuste V, et al. Medical management of connector pin thrombosis with the Amplatzer cardiac plug left atrial closure device. *World J Cardiol*. 2013;5:391-393.
11. López-Mínguez JR, Eldoayen-Gragera J, González-Fernández R, et al. Immediate and one-year results in 35 consecutive patients after closure of left atrial appendage with the Amplatzer Cardiac Plug. *Rev Esp Cardiol*. 2013;66:90-97.
12. Meincke F, Schmidt-Salzmann M, Kreidel F, et al. New technical and anticoagulation aspects for left atrial appendage closure using the Watchman device in patients not taking warfarin. *EuroIntervention*. 2013;9:463-468.
13. Chun KR, Bordignon S, Urban V, et al. Left atrial appendage closure followed by 6 weeks of antithrombotic therapy: A prospective single-center experience. *Heart Rhythm*. 2013;10:1792-1799.
14. Faustino A, Paiva L, Providência R, et al. Percutaneous closure of the left atrial appendage for prevention of thromboembolism in atrial fibrillation for patients with contraindication to or failure of oral anticoagulation: A single-center experience. *Rev Port Cardiol*. 2013;32:461-471.
15. Dilling-Boer D, Benit E, Herbots L, et al. Late organized left atrial thrombus on a left atrial appendage closure device. *J Cardiovasc Electrophysiol*. 2014;25:445-446.
16. Schiettekatte S, Czaplá J, Nijs J, et al. Unmasking a naked left atrial appendage closure device: a case of a silent embolic threat. *Heart Rhythm*. 2014;11:2314-2315.

17. Ternacle J, Lellouche N, Deux JF, et al. Left atrial appendage occluder thrombosis after successful implantation. *Circulation*. 2014;129:2576-2577.
18. Wiebe J, Bertog S, Franke J, et al. Safety of percutaneous left atrial appendage closure with the Amplatzer cardiac plug in patients with atrial fibrillation and contraindications to anticoagulation. *Catheter Cardiovasc Interv*. 2014;83:796-802.
19. Freixa X, Abualsaud A, Chan J, et al. Left atrial appendage occlusion: initial experience with the Amplatzer Amulet. *Int J Cardiol*. 2014;174:492-496.
20. Gafoor S, Franke J, Bertog S, et al. Left atrial appendage occlusion in octogenarians: short-term and 1-year follow-up. *Catheter Cardiovasc Interv*. 2014;83:805-810.
21. López-Mínguez JR, Asensio JM, Gragera JE, et al. Two-year clinical outcome from the Iberian registry patients after left atrial appendage closure. *Heart*. 2015;101:877-883.
22. Ketterer U, D'Ancona G, Siegel I, et al. Percutaneous left atrial appendage occlusion: Device thrombosis in clopidogrel non-responders. *Int J Cardiol*. 2016;204:196-197.
23. Kebernik J, Jose J, Abdel-Wahab M, et al. Safety and efficacy of left atrial appendage closure with the Amplatzer cardiac plug in very high stroke and bleeding risk patients with non-valvular atrial fibrillation. *Cardiol Ther*. 2015;4:167-177.
24. Freixa X, Scalone G, Martín-Yuste V, et al. Large protruding thrombus over left atrial appendage occlusion device successfully treated with apixaban. *Eur Heart J*. 2015;36:1427.
25. Świątkiewicz I, Woźnicki M, Sukiennik A, et al. Device-associated thrombus after percutaneous left atrial appendage closure: a Comunicación de caso and literature review. *Med Res J*. 2016;1:48-52.
26. Qazi AH, Wimmer AP, Huber KC, et al. Resolution (and late recurrence) of WATCHMAN device-related thrombus following treatment with dabigatran. *Echocardiography*. 2016;33:792-795.
27. Ledwoch J, Franke J, Gonzaga M, et al. Left atrial appendage closure: First in man with the 4th generation watchman device. *Catheter Cardiovasc Interv*. 2016;87:787-794.
28. Saw J, Tzikas A, Shakir S, et al. Incidence and clinical impact of device-associated thrombus and peri-device leak following left atrial appendage closure with the Amplatzer Cardiac Plug. *JACC Cardiovasc Interv*. 2017;10:391-399.